Message

Sent: 8/28/2019 8:02:30 PM

To: Volz, Stephanie [/o=ExchangeLabs/ou=Exchange Administrative Group

(FYDIBOHF23SPDLT)/cn=Recipients/cn=56abdbfdc7cc496ea71c6ac9ce20c36c-SVOLZ]

Subject: RE: PFAS Question

Attachments: strynar et al., 2015 SI.pdf; Strynar et al., 2015 ES&T.pdf

Stephanie,

First for the PFESA BP2, yes having two peaks is typical. Something about having asymmetric carbons or something like that. I don't recall the details. The peaks are usually exactly the same size, but for sure 2 peaks.

For the PFESA BP1 and the PFO5DoDA they are similar in molecular weight however two things will stand out.

- 1. The molecular weight between two compounds PFESA BP1 (442.9264) and PFO5DoDA (442.9442) is 40.2 ppm. Your QTOF should be able to tell them apart based on mass due to the EIC being extracted can be chosen to be +/- 5 ppm or so.
- 2. The PFESA BP1 will have a sulfonate (m/z 79.9) and a FSO3 (m/z 98.9) transition just like the PFOS, PFHxS and PFBS do.
- 3. PFO5DoDA will have a 84.9907 fragment however the PFESA BP1 will not. I have no idea what the 283.1367 fragment is.

Glad to chat if you need more later on tomorrow. Here is my 2015 paper in the SI that may have some details you can use.

Mark

From: Volz, Stephanie < Volz.Stephanie@epa.gov> Sent: Wednesday, August 28, 2019 3:06 PM To: Strynar, Mark < Strynar.Mark@epa.gov>

Subject: PFAS Question

Mark,

I just left you a phone message but was wondering if you have any insight or documentation into telling these two compounds (PFESA Byproduct 1 and PFO5DA) apart. I don't have any reference material but I am working on a QTOF.

Ex. 5 Deliberative Process (DP)

Ex. 5 Deliberative Process (DP)
on retention time or fragments for either compound.

I don't have background data

Compound Name Formula RT (min) Mass m/z Adduct/dimer Fragments Fragments Fragments CAS IUPAC N

PFESA (Byproduct 1) C7HF13O5S 443.93371 442.9441926 886.8596 (2M-H)-PFO5DA C7HF13O7 443.95147 442.926434 886.8951 (2M-H)- ? ? 29311-67-9 Perfluoro-3,? ? 39492-91-6 Perfluro-3,5

Also, I'm seeing a doublet for PFESA Byproduct 2 at 7.171 and 7.294 minutes. Do you know if that's typical?

Best Regards,

Stephanie Volz

US EPA/NEIC 1 Denver Federal Center, Bldg 25, Door E3 Denver, Colorado 80225 303-462-9127